

USER MANUAL

Steel 4-wheeled rollator

Product code: KORAL



1. Characteristics

The 4-wheeled, foldable rollator is made of a powder-coated steel frame. It features an upholstered, height-adjustable seat with a backrest and a convenient shopping bag. Rollator is equipped with height-adjustable handles with anatomical handgrips and brakes with a locking mechanism. It has narrow-profile wheels with a diameter of 8". The rollator is equipped with holders for a cane/crutch and a bottle.

The set includes: frame with seat (1 set), handles with brakes attached to the rear wheels (2 sets), front wheels (2 sets), backrest (1 pc.), mounting screws (6 pcs.), bottle holder (1 pc.)

Elements of the rollator:

- A. Handle
- B. Brake
- C. Mounting screw (wheel)
- D. Brake cables
- E. Backrest
- F. Seat
- G. Mounting screw (handles)
- H. Hinge of side crossbar
- I. Shopping bag
- J. Upper adjustment of brake
- K. Lower adjustment of brake
- L. Bottle holder



2. Application

The rollator is designed for patients with reduced mobility who need stable support and help while walking. It is also recommended to the elderly as an aid in keeping balance, improving walking comfort and ensuring safety from falling. The use of the rollator affects positively on the return to activity of the people

with locomotion and tilting to erect position problems. It guarantees better stabilization, support and security while walking.

3. Assembly and use

ASSEMBLY

- Unfold the rollator frame by laying both side hinges of the crossbar flat. Both crossbar hinges should be fully unfolded.
- Insert the left rear leg into the left rear opening of the frame. Lock the leg in place by tightening the mounting screw from the outside of the rollator frame.
- The left rear leg is connected to the left handle of the rollator via a brake cable. Insert the left handle into the left opening of the frame, adjusting it to the desired height. Route the brake cable outside the frame. Lock the handle at the correct height by tightening the mounting screw.
- Perform a similar installation for the right rear leg and the right handle, remembering to tighten the mounting screws.
- Insert the left front leg into the left front frame opening. Lock the leg height by tightening the mounting screw from the outside of the rollator.
- Perform a similar installation for the right front leg, remembering to tighten the mounting screw.

NOTE: All four legs should be set to the same height.

- Install the backrest by sliding it from the bottom up into the guides, located under the rollator handles.
- Install the bottle holder by sliding it from the top into the mounting bracket, located on the left side of the support frame.

UNFOLDING AND FOLDING THE ROLLATOR

- To fold the assembled rollator, lift the seat and pull the metal handle located above the bag. Then fold the rollator frame.
- To unfold the assembled rollator, unfold the frame by lying flat both side hinges of the crossbar located on the sides of the bag. Both crossbar hinges should be fully unfolded. Then lower the seat.

BRAKES FUNCTION

NOTE! When using the rollator for the first time, you may need assistance from another person, regardless of your physical condition.

- To stop the rollator, pull and hold the brake handle upwards.

- During a stop or while using the rollator seat, lock the brake by pushing the brake handle down.
- To release the brake, pull the brake handle upwards.

HEIGHT ADJUSTMENT OF HANDLES

The height of the handles should be adjusted so that the user can stand upright, as far as he or she is able to, with his or her shoulders in rest position and with slightly bent elbows. It is inadvisable to hang on the rollator with all the body weight.

To adjust the height of the handles, unscrew both mounting screws located on the outer side of the frame by, select appropriate height and tighten afterwards.

NOTE! Both handles must be set on the same level.

BRAKES ADJUSTMENT

- **Adjustment of the upper brake:** Unscrew the nut of the cable, pull it up to the appropriate length and lock it by tightening the locking screw.
- **Adjustment of the lower brake:** Loosen the mounting screw with the use of an open-end wrench, pull the black brake cable to adjust its tension, and then, tighten the mounting screw.

4. Technical data

Overall width	62 cm
Overall depth	58 cm
Seat dimensions	37 x 31 cm
Seat height from the ground	50,5-56 cm
Handles height adjustment	80,5-96 cm
Weight	8,6 kg
Maximum capacity	136 kg
Max. capacity of the shopping bag	5 kg
Frame colour	red

5. Notes

Timago International Group does not bear responsibility of improper use of the product, failure to comply safety regulations and misuse.

6. Maintenance

Clean the frame with damp cloth and dry thoroughly.

Do not use strong polishing agents due to the damage of protecting coatings.

Clean the dirt from moving parts of the rollator regularly (hinges, clamping screws, locks).

7. Storage and transport conditions

The product should not be exposed to direct sunlight, water or moisture.

8. Warranty information

All products distributed by our company are covered by a warranty, the terms of which are described in the warranty card available on our website. Please contact the retailer from whom you purchased the product. Please, keep in mind that for warranty purposes, the proof of purchase (receipt or invoice) should be retained.

9. Medical device incidents

Any serious incident involving a medical device must be reported to the manufacturer and to the competent authority of the European Union member state where the user or patient resides.

10. Labels

	Reference number		Manufacturer
	LOT number		UDI code
	Serial number		Medical device

	Note		Please read the instruction
	This product complies with the legal requirements of the European Union regarding medical devices (MDR 2017/745)		

Our complete line of products is available from our distributors. Learn more at: **www.timago.com**. Thank you for choosing Timago!