



designed for scientists



R 6547.1

/// Data Sheet

For safe positioning of the vessel below the bracket. A limit switch is attached to the container clamping device, which is connected to the electric control of the T 65 digital or RW 47 digital. The stirrer or disperser, which is mounted to the stand, can only be switched on when the vessel is fixed correctly by the clamping device.

Clamping device: textile strap, fastener made of stainless steel AISI 304.



designed for scientists

